

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 6/23/2014

DBH  
mmb

---

Instrument Serial Number: 010525 As of 23-Jun-14

---

Start Date	End Date	Agency Name
20-Jun-14		Chesapeake PD
11-Jun-14	20-Jun-14	DFS Central Lab
20-Jun-13	11-Jun-14	Montgomery County SO
12-Jun-13	20-Jun-13	DFS Central Lab
12-Dec-08	12-Jun-13	Montgomery County SO
27-May-08	12-Dec-08	DFS Central Lab

DBH  
mmB

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010525 Worksheet Start Date 6/20/2014  
Location Chesapeake PD 3rd  
Address 949 N George Washington Hwy Chesapeake VA 23323  
DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 759 Reference Barometer (mm HG) 761  
Reference Barometer(RB)Serial # 009109 RB Calibration Due 8/1/2014

## Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.304
0.300	0.300	0.291	0.309	Sample 2	0.303
Precision		sample min	sample max	Sample 3	0.303
0.001		0.303	0.304		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.100	0.097	0.103	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		

Dry gas standard Lot No. (with tank no.)

AG320502-21

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:

50

Notes:

Instrument Serial Number

010525

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

6/23/14

Issuing Analyst

Michelle Buller  
mdk

Date

6/26/14

Intox EC/IR-II: Accuracy Check

DBH  
mmB

*Department of Forensic Science*

Serial Number: 010525      Test Number: 1729  
Test Date: 06/20/2014      Test Time: 12:09 EDT  
Dry Gas Target: 0.300  
Lot Number: AG334603-07      Exp Date: 12/12/2015  
Tank Pressure: 223 psi      Barometric Pressure: 759 mmHg  
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:10
CHK	0.304	12:11
BLK	0.000	12:12
CHK	0.303	12:13
BLK	0.000	12:15
CHK	0.303	12:16

Test Status: *Success*

Calibration CRC: 776C41D7

Intox EC/IR-II: Accuracy Check

DBH  
mmB

*Department of Forensic Science*

Serial Number: 010525      Test Number: 1730

Test Date: 06/20/2014      Test Time: 12:19 EDT

Dry Gas Target: 0.100

Lot Number: AG320502-21      Exp Date: 07/24/2015

Tank Pressure: 776 psi      Barometric Pressure: 759 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:19
CHK	0.100	12:20
BLK	0.000	12:22
CHK	0.100	12:23
BLK	0.000	12:24
CHK	0.100	12:25

Test Status: *Success*

Calibration CRC: 776C41D7

DBH  
mmb




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED  INSTRUMENT, TEST,	NAME OF COURT  DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY  HALL, DONALD, B		AGENCY  DFS Central Lab
DFS LICENSE NUMBER  21250	LICENSE EXPIRES  10/01/2014	DATE TEST CONDUCTED  06/20/2014
TEST EQUIPMENT NUMBER  010525		

RESULTS: TIME SAMPLE TAKEN 12:33 EDT  
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20 \_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 20 Jun 2014 at 16:09

DBH  
mmB

## Test Results

Instrument Serial Number 010525

Test # 001731 Subject Test

Test Location 1 Department of  
Test Date 20 Jun 2014

Test Location 2 Forensic Science  
Test Time 12:27  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name HALL  
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B  
License Number 21250

Card Serial Number 121250  
Subject's Last Name INSTRUMENT

Effective Date 10/01/2012 Expiration Date 10/01/2014

Subject's Middle Initial

Subject's First Name TEST

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 20 Jun 2014 End Time 12:35

Result Time 12:33

Result Date 20 Jun 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:27

Data Type BLK

Sample Value 0.000

Sample Time 12:28

Data Type CHK

Sample Value 0.099

Sample Time 12:29

Data Type BLK

Sample Value 0.000

Sample Time 12:30

Data Type SUBJ

Sample Value 0.000

Sample Time 12:31

Data Type BLK

Sample Value 0.000

Sample Time 12:32

Data Type SUBJ

Sample Value 0.000

Sample Time 12:33

Data Type BLK

Sample Value 0.000

Sample Time 12:34

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG320502-21

Standard Expiration Date 07/24/2015

Tank Pressure 772

Barometric Pressure 759 mmHg

Blow Sample Number 1 Blow Duration 3.84 sec

Blow Volume 1963 cc End-of-Blow Time 12:31

Blow Sample Number 2 Blow Duration 4.00 sec

Blow Volume 2157 cc End-of-Blow Time 12:33

Tamper Evident Stamp d68970bc

Test Status Code 0

Test Status Success